



The longer length of Philips AngioSculpt provides more coverage and convenience for infrainguinal artery procedures, but it's the unique advantage of the AngioSculpt scoring technology that really delivers.

Number	Balloon diameter (mm)	Balloon length (mm)	Catheter length (cm)	Guidewire compatibility	Sheath size (F)
2039-2010	2.0	10	137	0.014"	5F
2039-2020	2.0	20	137	0.014"	5F
2155-2040	2.0	40	155	0.014"	5F
2215-20100	2.0	100	137	0.014"	6F
2216-20100	2.0	100	155	0.014"	6F
2039-2520	2.5	20	137	0.014"	5F
2155-2540	2.5	40	155	0.014"	5F
2215-25100	2.5	100	137	0.014"	6F
2216-25100	2.5	100	155	0.014"	6F
2039-3020	3.0	20	137	0.014"	5F
2155-3040	3.0	40	155	0.014"	5F
2215-30100	3.0	100	137	0.014"	6F
2216-30100	3.0	100	155	0.014"	5F
2039-3520	3.5	20	137	0.014"	5F
2155-3540	3.5	40	155	0.014"	5F
2215-35100	3.5	100	137	0.014"	6F
2216-35100	3.5	100	155	0.014"	6F
2076-4020	4.0	20	137	0.018"	6F
2092-4040	4.0	40	90	0.018"	6F
2076-4040	4.0	40	137	0.018"	6F
2290-40100	4.0	100	90	0.014"	6F
2237-40100	4.0	100	137	0.014"	6F
2249-40200	4.0	200	137	0.014"	6F

Number	Balloon diameter (mm)	Balloon length (mm)	Catheter length (cm)	Guidewire compatibility	Sheath size (F)
2249-40200	4.0	200	137	0.014"	6F
2076-5020	5.0	20	137	0.018"	6F
2092-5040	5.0	40	90	0.018"	6F
2076-5040	5.0	40	137	0.018"	6F
2290-50100	5.0	100	90	0.014"	6F
2237-50100	5.0	100	137	0.014"	6F
2249-50200	5.0	200	137	0.014"	6F
2105-6020	6.0	20	50	0.018"	6F
2092-6020	6.0	20	90	0.018"	6F
2076-6020	6.0	20	137	0.018"	6F
2105-6040	6.0	40	50	0.018"	6F
2092-6040	6.0	40	90	0.018"	6F
2076-6040	6.0	40	137	0.018"	6F
2290-60100	6.0	100	90	0.014"	6F
2237-60100	6.0	100	137	0.014"	6F
2249-60200	6.0	200	137	0.014"	6F
2332-7040	7.0	40	50	0.018"	6F
2333-7040	7.0	40	90	0.018"	6F
2334-7040	7.0	40	137	0.018"	6F
2332-8040	8.0	40	50	0.018"	6F
2333-8040	8.0	40	90	0.018"	6F
2334-8040	8.0	40	137	0.018"	6F

Important safety information

Indications

The AngioSculpt PTA scoring balloon catheter is intended for dilatation of lesions in the iliac, femoral, ilio-femoral, popliteal, infra-popliteal, and renal arteries, and for the treatment of obstructive lesions of native or synthetic arteriovenous dialysis fistulae. Not for use in the coronary or neuro-vasculature.

Contraindications

None known for percutaneous transluminal angioplasty (PTA) procedures.

Warnings

This device is intended for single (one) patient use only. Do not resterilize and/or reuse, as this can potentially result in compromised device performance and increased risk of inappropriate resterilization and cross contamination. The inflated diameter of the balloon should approximate the diameter of the vessel just proximal and distal to the stenosis, in order to reduce potential vessel damage. When the catheter is exposed to the vascular system, it should be manipulated while under high-quality fluoroscopic observation. Do not advance or retract the catheter unless the balloon is fully deflated under vacuum. If resistance is met during manipulation, determine the cause of the resistance before proceeding. Balloon pressure should not exceed the rated burst pressure (RBP). Refer to product label for device-specific information. The RBP is based on results of in-vitro testing. At least 99.9% of the balloons (with a 95% confidence level) will not burst at or below their RBP. Use of a pressure monitoring device is recommended to prevent over-pressurization. Use only the recommended balloon inflation medium. Never use air or any gaseous medium to inflate the balloon. Proceed cautiously when using the AngioSculpt catheter in a freshly deployed bare metal or drug-eluting stent. The AngioSculpt catheter has not been tested for post-dilation of stents or in lesions distal to freshly deployed stents in clinical studies. Bench testing has shown no additional risk when inserting or withdrawing the AngioSculpt catheter through stents (no interference with stent struts, no retention of or damage to the AngioSculpt catheter). Use the catheter prior to the "Use before" (expiration) date specified on the package.

Precautions

A thorough understanding of the principles, clinical applications and risks associated with PTA is necessary before using this product.

Any use for procedures other than those indicated in these instructions is not recommended. The device is not recommended for use in lesions that may require inflation pressures higher than those recommended for this catheter. Do not use if package is opened or damaged. Prior to angioplasty, the catheter should be examined to verify functionality, device integrity and to ensure that its size and length are suitable for the specific procedure for which it is to be used. During and after the procedure, appropriate anticoagulants, antiplatelet agents and vasodilators should be administered to the patient according to institutional practice for peripheral angioplasty of similar arteries. Pass the AngioSculpt catheter through the recommended introducer sheath size or minimum size guiding catheter indicated on the product label.

Adverse effects

Possible adverse effects include, but are not limited to, total occlusion of the treated artery, arterial dissection or perforation, arterial spasm, pseudoaneurysm, restenosis of the dilated artery, embolism, thrombus, retained device components, hemorrhage or hematoma, arteriovenous fistula.

